differential response to compression at various depths from the articular surface to the tide line in normal canine femoral condylar cartilage. To do so they employed 100 μm slices mounted on a transmitted light Nomarski microscope in a stage mounted loading rig. In these essentials the method is unchanged, but it has been subject to improvements in three ways, the subject of the present report.

1. Acquisition of strain data We sought a method of acquiring data quickly and concluded that a computer based system would be effective. We have therefore coupled a video image processor (VIP, Sight Systems, Berkshire) to a BBC microcomputer. For each analysis we record the image of compressing cartilage with a video camera (RCA TC 1004) and VHS tape recorder (Panasonic NV-180). We make measurements during tape playback using a modified software program, 'U. locate' (Sight Systems). This is devised so that the cursor scans a given image area for 'peak black' and hence follows any movement of that image. Chondrocytes are suitable peak black images; they move during cartilage compression and can thus be tracked, one per tape playback. Successively deeper chondrocytes are selected in a broadly similar pattern to that adopted by O'Connor *et al.* During a tracking sequence at set time intervals—for example 1.5 s—the position of the chondrocyte is marked on the screen by a stationary cross. By the end of each tape playback the screen carries a number of these crosses; their positions are automatically stored in the computer memory and their x, y coordinates are displayed and printed out. These measurements of chondrocyte displacement enable the compressibility of the cartilage throughout its depth to be rapidly gauged.

2. Tissue constraint chamber A haemocytometer has clear advantages over chambers employed by O'Connor *et al.* Their chambers were fabricated individually for each test slice of cartilage from slide, cover-slip, shim steel spacers, and glue. The haemocytometer has improved characteristics of forming a reusable chamber 100 μm deep, for which a standard anvil and backplate pair are manufactured.

3. Advance of the anvil A sustained known rate of advance of the anvil was required hence manual operation has been superseded by motor drive using a 15 volt DC motor controlled by a square wave generator. We employ a 'control-a-train' kit (LK64U, Maplin Electronics, UK) and rates of anvil advance of up to 50 $\mu\text{m/s}$.

The improved micromechanical test apparatus is robust, reliable, relatively simple to operate, and not unduly expensive. We believe that the described advances will facilitate understanding of the behaviour of articular cartilage during the loading cycle.

Department of Histopathology,
University Hospital of
South Manchester
Nell Lane,
Withington,
Manchester M20 8LR

M WARSKYJ
P SULLIVAN
P O'CONNOR
D M LAWTON
D L GARDNER

Reference

- 1 O'Connor P, Orford C R, Gardner D L. Differential response to compressive loads of zones of canine hyaline articular cartilage: micromechanical, light and electron microscopic studies. *Ann Rheum Dis* 1988; 47: 414–20.

Book reviews

Rheumatology in Clinical Practice. By J M H Moll. Pp. 764. £69.50. Blackwell: Oxford. 1987. ISBN: 0-632-01342-7.

The author aims, 'to bridge the gap between introductory texts and detailed monographs'. The book has 764 pages and is divided into six sections. Section I highlights milestones in rheumatological history and makes fascinating reading. Section II summarises the scientific basis of rheumatology in conventional fashion with chapters devoted to anatomy, physiology, biochemistry, pathology, immunology, and immunogenetics. Section III covers clinical, radiological, and laboratory assessment and section IV describes the spectrum of rheumatic disorders with separate chapters for rheumatoid arthritis, osteoarthritis, spondyloarthritides, crystal deposition disorders, connective tissue disorders, rheumatic disorders of childhood, back pain, soft tissue rheumatism, and a miscellany of other conditions. Section V discusses management in the broadest sense with a thought provoking chapter on 'Communication', and

Section VI gives a personal overview discussing such questions as: 'What is rheumatology?'; 'What is the size and impact of the rheumatological problem?'; 'How will future research add to rheumatological knowledge?'.

The layout is clear with an index of topics at the start of each chapter and plenty of references. I particularly liked the short summaries that emphasise the 'take home message' at the end of chapters. The book is notable for its widespread use of line and halftone drawings by the author, though good use is also made of clinical photographs.

Rheumatology in Clinical Practice is comprehensive but not fussy, easy to read, and stamped with the personality of the author. It offers an alternative to the standard textbooks at the price of £69.50.

Bone and Joint Research Unit,
The London Hospital Medical College,
Turner Street,
London E1 2AD

PAUL W THOMPSON