Conclusion: This study shows the high prevalence of echocardiographic alterations in PsA patients compared to the general population, of the same magnitude as patients with RA. We emphasize the value of an echocardiogram for a complete cardiovascular evaluation and early detection of cardiac abnormalities in these patients.

REFERENCES:

- [1] Shang Q, Tam LS, Yip GW, et al. High prevalence of subclinical left ventricular dysfunction in patients with psoriatic arthritis. *J Rheumatol* 2011;38(7):1363-70. doi: 10.3899/jrheum.101136

Disclosure of Interests: None declared

DOI: 10.1136/annrheumdis-2021-eular.1439

POS1393

QUANTITATIVE AUTOFLUORESCENCE FINDINGS IN PATIENTS UNDERGOING HYDROXYCHLOROQUINE TREATMENT

S. Parrulli¹, M. Cozzi¹, M. Airal¹, F. Romano¹, F. Viola², P. Sarzi-Puttini³, G. Staurengi¹, A. Invernizzi^{1,4}, ¹Sacco Hospital, University of Milan, Eye Clinic, Department of Biomedical and Clinical Science "Luigi Sacco", Milan, Italy; ²University of Milan, IRCCS-Ca Granda Fondazione Ospedale Maggiore Policlinico, Ophthalmological Unit, Department of Clinical Sciences and Community Health, Milan, Italy; ³ASST Fatebenefratelli L. Sacco University Hospital, Rheumatology Unit, Milan, Italy; ⁴Sydney Medical School, The University of Sydney, Save Sight Institute, Discipline of Ophthalmology, Sydney, Australia

Background: Hydroxychloroquine (HCQ) is a relatively safe and effective drug widely used as primary or adjunctive treatment for several rheumatological and dermatological disorders¹. HCQ modulates immune response through several mechanisms and has a tropism for pigmented ocular tissues, particularly retinal pigment epithelium (RPE)². Its accumulation within RPE cells can lead to sight threatening retinal toxicity, with bull's eye maculopathy (BEM) representing its advanced phenotype.³ Quantitative Auto-Fluorescence (qAF) is an imaging modality that allows the measurement of retinal auto-fluorescence following short-wavelength light (488nm) excitation of retinal fluorophores (lipofuscin).⁴ Two recent studies have focused on qAF values in patients treated with HCQ^{5,6}. In both cases qAF was increased in eyes with BEM. Furthermore, Reichel et al.⁶ were able to detect increased values of qAF in patients without BEM as early as 6 months after the start of HCQ treatment using an experimental imaging analysis procedure.

Objectives: To measure quantitative autofluorescence (qAF) in patients under treatment with hydroxychloroquine (HCQ) with no apparent signs of retinal toxicity and to compare it with that of untreated subjects.

Methods: Consecutive patients at risk for the development of HCQ retinal toxicity (duration of treatment >5 years or daily HCQ dose >5 mg/kg of actual body weight (ABW) and/or renal insufficiency)⁷ but no alterations on Spectral Domain - Optical Coherence Tomography, Short-Wavelength Autofluorescence and 10-2 Visual Field examination were recruited. Healthy subject matched by age and sex were also enrolled in the study. All subjects underwent qAF measurements in one eye. Images were analyzed using the conventional qAF grid by Delori calculating the qAF of 8 sectors of the intermediate ring and the mean of those values (qAF₈).

Results: Thirty-nine patients treated with HCQ (38 females, mean age 52,1 ± 8,6 years) and 39 untreated subjects (38 females, mean age 51,2 ± 8,6 years). In both HCQ patients and untreated subjects, qAF₈ was positively correlated with age (p=0.004) (Figure 1). Although HCQ patients showed a higher mean qAF₈ compared to untreated subjects (294,7 ± 65,3 vs 268,9 ± 57,5), the difference was not significant (p=0.068). HCQ patients showed significantly higher mean qAF values in the inferior-temporal, inferior and inferior-nasal sectors of the intermediate ring of qAF grid compared to untreated subjects (all p<0.05).

Figure 1. Visual representation of a model predicting the standardized qAF values as influenced by age and HCQ daily dose/ABW, calculated for a treatment duration of 15 years.

Conclusion: These results suggest a possible preclinical increase of qAF values in inferior parafoveal sectors probably induced by HCQ exposure. Further studies are required to improve our understanding of preclinical stages of HCQ retinopathy and the possible role of qAF in the HCQ toxicity screening.

REFERENCES:

- [1] Haładyj, E., Sikora, M., Felis-Giemza, A. & Olesińska, M. Antimalarials - are they effective and safe in rheumatic diseases? *Rheumatologia* 56, 164–173 (2018).
- [2] Rosenthal, A. R., Kolb, H., Bergsma, D., Huxsoll, D. & Hopkins, J. L. Chloroquine retinopathy in the rhesus monkey. *Invest. Ophthalmol. Vis. Sci.* 17, 1158–1175 (1978).
- [3] Modi, Y. S. & Singh, R. P. Bull's-Eye Maculopathy Associated with Hydroxychloroquine. *N. Engl. J. Med.* 380, 1656 (2019).
- [4] Sparrow, J. R., Duncker, T., Schuerch, K., Paavo, M. & de Carvalho, J. R. L. J. Lessons learned from quantitative fundus autofluorescence. *Prog. Retin. Eye Res.* 74, 100774 (2020).
- [5] Greenstein, V. C. et al. Quantitative Fundus Autofluorescence in HCQ Retinopathy. *Invest. Ophthalmol. Vis. Sci.* 61, 41 (2020).
- [6] Reichel, C. et al. Quantitative Fundus Autofluorescence in Systemic Chloroquine/Hydroxychloroquine Therapy. *Transl. Vis. Sci. Technol.* 9, 42 (2020).
- [7] Yusuf, I. H., Sharma, S., Luqmani, R. & Downes, S. M. Hydroxychloroquine retinopathy. *Eye (Lond)* 31, 828–845 (2017).

Disclosure of Interests: Salvatore Parrulli: None declared, Mariano Cozzi Grant/research support from: Bayer, Nidek, Zeiss, Matteo Airal: None declared, Francesco Romano: None declared, Francesco Viola: None declared, Piercarlo Sarzi-Puttini: None declared, Giovanni Staurengi Grant/research support from: Heidelberg Engineering (C), QuantelMedical (C), Centervue (C), Carl Zeiss Meditec (C), Alcon (C), Allergan (C), Bayer (C), Boehringer (C), Genentech (C), GSK (C), Novartis (C), and Roche (C), Optos (F), Optovue (F) and Centervue (F), Alessandro Invernizzi Grant/research support from: Novartis (C), Bayer (C)

DOI: 10.1136/annrheumdis-2021-eular.2022

POS1394

ACCURACY AND PERFORMANCE OF A HANDHELD ULTRASOUND DEVICE TO ASSESS ARTICULAR AND PERIARTICULAR PATHOLOGIES IN PATIENTS WITH INFLAMMATORY ARTHRITIS

G. Corte^{1,2}, S. Bayat^{1,2}, K. Tascilar^{1,2}, L. Valor^{1,2}, L. Schuster^{1,2}, J. Knitz^{1,2}, G. Schett^{1,2}, A. Kleyer^{1,2}, D. Simon^{1,2} on behalf of <https://www.medin3n3.uk-erlangen.de/en/research/research-main-focus/digital-rheumatology-and-experimental-imaging/>. ¹Friedrich-Alexander-Universität (FAU) and Universitätsklinikum Erlangen, Internal Medicine 3, Rheumatology and Immunology, Erlangen, Germany; ²Friedrich-Alexander University (FAU) Erlangen-Nürnberg and Universitätsklinikum Erlangen, Deutsches Zentrum Immuntherapie, Erlangen, Germany

Background: Handheld ultrasound (HHUS) devices have increasingly found their way into clinical practice due to several advantages (e.g. portability, significantly lower purchase cost). However, there is no evidence to date on the accuracy and performance of HHUS in patients with inflammatory arthritis (IA).

Objectives: To assess accuracy and performance of a new HHUS machine in comparison to a conventional cart-based sonographic machine in patients with IA.

Methods: Consecutive IA patients of our outpatient clinic with at least one tender and swollen joint in the 66/68 joint count were enrolled. US was performed on clinically affected joints with corresponding tendons/enthese using a cart-based sonographic device ("Samsung HS40") and a HHUS device ("Butterfly iQ") in standard scan positions. One blinded reader scored all images for the presence of following pathologic findings: erosions, bony enlargement, synovial hypertrophy, joint effusion, bursitis, tenosynovitis and enthesitis. In addition, synovitis was graded (B Mode and power Doppler (PD)) by the 4-level EULAR-OMERACT scale [1]. To avoid bias by the blinded reader, who otherwise would have been tempted to identify pathologic findings for each examined joint, we also included 67 joints of two healthy volunteers into the evaluation. We calculated the overall concordance and the concordance by type of joint and type of pathological finding between the two devices (percentage of observation pairs in which the same rating was given by both devices). The Cohen's kappa coefficient (κ) with 95% bootstrap confidence intervals was used to assess the agreement between the two US devices. We also measured the time required for the US examination of one joint with both devices.

Results: 32 patients (20 rheumatoid arthritis, 10 psoriatic arthritis, 1 gouty arthritis, 1 systemic lupus erythematosus) were included in this study. Mean age of patients was 58.2±13.7 years, 63% were females. In total 186 joints were examined. The overall raw concordance in B-mode between the two devices was 97 %, with an overall kappa for agreement of 0.90, 95% CI (0.89, 0.94). No significant differences were found in relation to type of joint or pathological finding examined. The PD-mode of the HHUS device did not detect any PD-signal, whereas the cart-based device detected a PD-signal in 61 joints (33%). The portable device did not offer any time saving compared to the cart-based device (mean time in seconds per examined region: 47 seconds for the HHUS device versus 46.3 seconds for the cart-based device).

Conclusion: The HHUS device "Butterfly iQ" has been shown to be accurate in the assessment of structural joint damage and inflammation in patients with IA,